






Category	Product Image	Product #	INFOR #	Product Name	Manufacturer Name	Manufacturer Product #	Filtration	Fluid Resistance	Surgical Use Acceptable	Non Procedural Non Surgical Acceptable	CDC/NIOSH Approval Number	CDC/NIOSH/FDA Approved	Location Deployment as of 11/2/20	
F I T T E S T E D F F R - N 9 5	Conventional													
		MMM1860	218589	MASK,RESP,N95,NIOSH APPROVED,MEDIUM	3M HEALTHCARE	1860	N95	120mmHG	YES ASTM Level 2	YES	YES 84A-0006	CDC / NIOSH approved	FIT test locations per PPE policy	
		MMM1860S	218590	MASK,RESP,N95,NIOSH APPROVED,SMALL	3M HEALTHCARE	1860S	N95	120mmHG	YES ASTM Level 2	YES	YES 84A-0006	CDC / NIOSH approved	FIT test locations per PPE policy	
		MMM1870	161396	MASK,AURA,SURG,PARTICULATE, PLUS,N95	3M HEALTHCARE	1870+	N95	160mmHG	YES ASTM Level 3	YES	YES 84A-3844	CDC / NIOSH approved	FIT test locations per PPE policy	
		MMM9205PB	221410	Mask, Aura, Particulate Respirator 9205+	3M Healthcare	9205+	N95	NA	Y	Yes	Yes TC-94A-8590	CDC / NIOSH approved	FIT test locations per PPE policy	
		MOM1510	234082	RESPIRATOR,W/MASK, X-SM,N95,LF,20/BX	MOLDEX-METRIC INC	1510	N95	160mmHG	YES ASTM Level 3	YES	YES 84A-5171	CDC / NIOSH approved	FIT test locations per PPE policy	
	Contingent													
		8210R	219760	MASK RESPIRATOR 3M N95	3M HEALTHCARE	8210R	N95	NA	Yes with overlying surgical facemask or faceshield	YES	YES 84A-0007	CDC / NIOSH approved	FIT test locations per PPE policy	
		8210PLUS	220389	MASK RESPIRATOR 3M N95 PLUS	3M HEALTHCARE	8210P	N95	NA	Yes with overlying surgical facemask or faceshield	YES	YES 84A-0007	CDC / NIOSH approved	FIT test locations per PPE policy	
		8110S	220992	MASK RESPIRATOR 3M N95	3M HEALTHCARE	8110S	N95	NA	Yes with overlying surgical facemask or faceshield	YES	TC-84A-0007	CDC / NIOSH approved	FIT test locations per PPE policy	
		9210P+	221411	MASK RESPIRATOR 3M N95	3M HEALTHCARE	9210+	N96	NA	Yes with overlying surgical facemask or faceshield	YES	TC-84A-5668	CDC / NIOSH approved	REPLACEMENT FOR 1870 - INDUSTRIAL	
		9205+	221410	MASK RESPIRATOR 3M N95	3M HEALTHCARE	9205+	N97	NA	Yes with overlying surgical facemask or faceshield	YES	TC-84A-8590	CDC / NIOSH approved	REPLACEMENT FOR 1870 - INDUSTRIAL	

Category	Product Image	Product #	INFOR #	Product Name	Manufacturer Name	Manufacturer Product #	Filtration	Fluid Resistance	Surgical Use Acceptable	Non Procedural Non Surgical Acceptable	CDC/NIOSH Approval Number	CDC/NIOSH/FDA Approved	Location Deployment as of 11/2/20
NON FIT TESTED FFR		8210 VALVE	220394	MASK RESPIRATOR 3M N95 VENT/INDUSTRIAL	3M HEALTHCARE	8210V	N95	NA	Yes with overlying surgical facemask	Yes with overlying facemask	YES 84A-5410	CDC / NIOSH approved	Any N95 to be deployed at any Network
		1804S	230993	MASK RESPIRATOR 3M N95 Surgical	3M HEALTHCARE	1804S	N95	80 mmHg	YES With Face Shield ASTM Level 1	YES	TC-84A-7790	CDC / NIOSH/FDA approved	Any N95 to be deployed at any Network
		8000	220196	MASK RESPIRATOR 3M N95 8000	3M HEALTHCARE	8000	N95	NA	Yes with overlying surgical facemask or faceshield	YES	YES 84A-3981	CDC / NIOSH approved	Any N95 to be deployed at any Network
		8511 VALVE	220388	MASK RESP N95 3M REG N-LATEX	3M HEALTHCARE	8511	N95	INDUSTRY	Yes with overlying surgical facemask	Yes with overlying facemask	YES 84A-1299	CDC / NIOSH approved	Any N95 to be deployed at any Network
		JFY	220222	MASK RESPIRATOR	JFY	4150	N95	not specified	Yes with overlying surgical facemask or faceshield	YES	YES 84A-3888	CDC / NIOSH approved	Any N95 to be deployed at any Network
		9500-N95	220260	MASK RESPIRATOR MAKRITE N95 REG	MAKRITE	9500-N95	N95	not specified	YES	YES	YES 84A-5411	CDC / NIOSH approved	Any N95 to be deployed at any Network
		N95- S	220301	MASK RESP N95 N-LATEX SMALL	CARDINAL HEALTHCARE	1510	N95	not specified	YES	YES	YES 84A-3323	CDC / NIOSH approved	Any N95 to be deployed at any Network
		N95- M	220302	MASK RESP N95 N-LATEX IN SIZE MEDIUM ONLY	CARDINAL HEALTHCARE	1510	N95	not specified	YES	YES	YES 84A-3323	CDC / NIOSH approved	Any N95 to be deployed at any Network
		SH9550	220950	Mask	San Huei United Company, Ltd.	SH9550	N-95	not specified	Yes with overlying surgical facemask or faceshield	YES	YES 84A-3713	CDC / NIOSH approved	Any N95 to be deployed at any Network
		46767	121259	MASK RESP N95 TECHNOL REGULAR	HALYARD HEALTH INC	46767	N95	160mmHG	YES ASTM Level 3	YES	YES 84A-7523	CDC / NIOSH approved	Any N95 to be deployed at any Network
		46867	121258	MASK RESP N95 TECHNOL SMALL	HALYARD HEALTH INC	46867	N95	160mmHG	YES ASTM Level 3	YES	YES 84A-7520	CDC / NIOSH approved	Any N95 to be deployed at any Network
	H950	220575	Non-Valve Respirator N95	Home Depot	H950	N-95	Not specified	Yes with overlying surgical facemask or face shield	YES	YES 84A-5411	CDC / NIOSH approved	Any N95 to be deployed at any Network	

Category	Product Image	Product #	INFOR #	Product Name	Manufacturer Name	Manufacturer Product #	Filtration	Fluid Resistance	Surgical Use Acceptable	Non Procedural Non Surgical Acceptable	CDC/NIOSH Approval Number	CDC/NIOSH/FDA Approved	Location Deployment as of 11/2/20
		46827	221291	N95 Particulate Filter Respirator and Surgical Mask	HALYARD HEALTH INC	46827	N95	160mmHG	YES ASTM Level 3	YES	YES 84A-7518	CDC / NIOSH approved	Any N95 to be deployed at any Network
		N1105M/L	222679	Non-Valve Respirator N95	Sperian - Honeywell	N1105M/L	N-95	not specified	Yes with overlying surgical facemask or face shield	YES	YES 84A-3789	CDC / NIOSH approved	Any N95 to be deployed at any Network